



EU DECLARATION OF CONFORMITY

Manufacturer Name/Address: Ansell Healthcare Europe NV
Boulevard International 55
Brussels
B-1070
Belgium

SRN Number: BE-MF-000000691

Risk Class: Class I

Intended Purpose: A non-sterile medical device intended as an examination/treatment glove and a protective barrier when worn on the hands of healthcare providers, during patient examination, or for other sanitary purposes. The device is used mainly as a two-way barrier to protect both the patient and wearer against contaminants. This is a single-use device.

EMDN Code and Description: T01020204 - Nitrile Examination/Treatment Glove

Basic UDI-DI: 5414566 MF93852 EX

Product Name(s):

Product Name	Size	Product Code	Market Regions
MICROFLEX® 93-852	XS	93852060	EMEA, APAC
MICROFLEX® 93-852	S	93852070	EMEA, APAC
MICROFLEX® 93-852	M	93852080	EMEA, APAC
MICROFLEX® 93-852	L	93852090	EMEA, APAC
MICROFLEX® 93-852	XL	93852100	EMEA, APAC
MICROFLEX® 93-852	XXL	93852110	EMEA, APAC

Conformity Assessment Procedure: Annex I & Annex II + Annex III

This EU Declaration of Conformity is issued under the sole responsibility of Ansell Healthcare Europe NV.

Ansell Healthcare Europe NV/SA

—
Riverside Business Park – Block J
Boulevard International 55
1070 Brussels, Belgium

T. + 32 (0)2 528 74 00
F. + 32 (0)2 528 74 01
www.ansell.com



We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

Signed on behalf of Ansell Healthcare Europe NV

A handwritten signature in black ink, appearing to read "S. Marshall", is written over a light blue rectangular background.

Ansell Healthcare Europe NV
Riverside Business Park - Block J
Bld Internationaalelaan 55
B-1070 Brussels
BELGIUM

Name: Samantha Marshall
Position: Director Regulatory Affairs Medical
Date of issue: 16 December 2024
Place of issue: Nuneaton, England
Version No: MED\MFX93852\005