

High-visibility orange nitrile exam gloves, for protection that you can see

- **High visibility:** MICROFLEX™ 93-856 are robust, high visibility orange nitrile medical exam gloves that protect hands in demanding environments — their high-visibility orange color makes them easy to see in emergency situations
- **Secure grip:** These gloves work exceptionally well in oily environments — with their textured fingers, wearers can confidently grip tools and materials in every situation
- **Added protection:** These orange nitrile gloves also feature an extended cuff for added protection of the wrist and forearm, and offer additional protection against chemotherapy drugs*



*see product packaging for details

Industries

- Automotive
- Automotive Aftermarket
- Life Sciences
- Mining
- Energy
- Healthcare

Recommended For

- Inspection, selecting, checking parts
- Assembly and inspection of components
- Equipment repair and maintenance
- General Purpose Auto Aftermarket
- Sample taking and processing
- Administering drugs
- Raw material sample collection
- Picking parts and accessories
- Inspection of primary product
- Product inspection
- Vehicle maintenance
- Handling of tools
- Extra protection over the wrist and arm
- Protection from Type I latex allergy in HCW's or patients

TECHNICAL DATA SHEET

| Product Information | |
|-------------------------------|--|
| Material | Nitrile |
| External Glove Surface | Textured Fingers |
| Audit Standards | ISO 13485:2003 |
| Product Certification | Medical Device Regulation (EU) 2017/745 Class I, Personal Protective Equipment Regulation (EU) 2016/425 Category III risks, CE 0493, European Directive 89/686/EEC, EN 455 1-4, EN ISO 374-1:2016 Type B JKT, EN ISO 374-5:2016 Virus, FDA21 CFR 177-2600-US Food Contact Approved, UKCA |
| Packaging Overview | 100 gloves per dispenser 10 dispensers per case 1000 gloves per case |
| Storage Instructions | Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition. |
| Country Of Origin | Malaysia |
| Product Segmentation | High Risk |
| Antistatic | Yes EN1149 |
| Silicone Free | No |

PHYSICAL PROPERTIES

| | Typical Values | | Testing Method |
|---|----------------|-------------|---------------------|
| | BEFORE AGING | AFTER AGING | |
| Length (mm/inches) | 270 / 10.6 | | ASTM D3767,EN 420 |
| Freedom from Holes (Inspection level I) | 1.5 AQL | | ASTM D5151,EN 455-1 |
| Palm Thickness (mm/mils) | 0.13 / 5.1 | | ASTM D3767,EN 420 |
| Finger Thickness (mm/mils) | 0.20 / 7.9 | | ASTM D3767,EN 420 |
| Elongation at Break (%) | ≥ 500 | ≥ 400 | ASTM D412 |
| Force at break (N) | ≥9 | ≥10 | EN 455-2 |

ORDERING INFORMATION

| Size | S (6.5 - 7) | M (7.5 - 8) | L (8.5 - 9) | XL (9.5 - 10) | XXL (10.5 - 11) |
|--------------|-------------|-------------|-------------|---------------|-----------------|
| Product Code | 93856070 | 93856080 | 93856090 | 93856100 | 93856110 |

Performance Standards and Regulatory Compliance



For additional information visit us at www.ansell.com, or call us at

Europe, Middle East & Africa Region

Ansell Healthcare Europe NV
T: +32 (0) 2 528 74 00
F: +32 (0) 2 528 74 01

Asia Pacific Region

MY: +603 8310 6688
apac.medical@ansell.com
CN: +86 21 38275000
infochina@ansell.com
JP: +813 5549 8151
info.medical.jp@ansell.com

North America Region

Ansell Healthcare Products LLC
T: +1 800 800 0444
F: +1 800 800 0445

Latin America & Caribbean Region

BR: +55 (11) 3356-3100
latam.medical@ansell.com
MX: +52 442 296 2050
latam.medical@ansell.com
CO: +57 1 288 3247
latam.medical@ansell.com

Australia

Ansell Limited
T: +61 1800 337 041
F: +61 1800 803 578

Russia

Ansell PYC
Ten. +7 495 258 13 16

Ansell, ® and ™ are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending; www.ansell.com/patentmarking © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

