The Manufacturer
ANSELL HEALTHCARE EUROPE N.V.
RIVERSIDE BUSINESS PARK, BLOCK J
BOULEVARD INTERNATIONAL 55
B-1070 BRUSSELS
BELGIUM

declares under his sole responsibility, that the PPE described hereafter:

## MicroFlex® NeoTouch 25-101

Products manufactured as of: [2024/11/27]

PPE to be used against category III risks

EN ISO 374-1:2016 Type B

EN ISO 374-5 VIRUS

is in conformity with the provisions of Regulation (EU) 2016/425 and with the standards EN ISO 21420:2020, EN ISO 374-1:2016, EN ISO 374-5:2016 and is identical to the PPE which is subject to the EU-Type examination (Module B, Annex V of the Regulation), under certificate number 032/2021/1313.02, issued by the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

and is subject to the procedure set out in Annex VIII (Module D) of the Regulation under the supervision of the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

Guido Van Duren

Director - Regulatory affairs

Ansell

Place: Brussels Date: 2024/11/27

The Manufacturer
ANSELL HEALTHCARE EUROPE N.V.
RIVERSIDE BUSINESS PARK, BLOCK J
BOULEVARD INTERNATIONAL 55
B-1070 BRUSSELS
BELGIUM

declares under his sole responsibility, that the PPE described hereafter:

# MicroFlex® NeoTouch 25-101

Products manufactured as of: [2021/12/03] and till: [2024/11/26]

PPE to be used against category III risks





is in conformity with the provisions of Regulation (EU) 2016/425 and with the standards EN ISO 21420:2020, EN ISO 374-1:2016, EN ISO 374-5:2016 and is identical to the PPE which is subject to the EU-Type examination (Module B, Annex V of the Regulation), under certificate number 032/2021/1313, issued by the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

and is subject to the procedure set out in Annex VII (Module C2) of the Regulation under the supervision of the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

Guido Van Duren Director - Regulator

Director - Regulatory affairs

Ansell

Place: Brussels Date: 2021/12/03

The Manufacturer
ANSELL HEALTHCARE EUROPE N.V.
RIVERSIDE BUSINESS PARK, BLOCK J
BOULEVARD INTERNATIONAL 55
B-1070 BRUSSELS
BELGIUM

declares under his sole responsibility, that the PPE described hereafter:

## MicroFlex® NeoTouch 25-101

Products manufactured as of: [2018/04/21] and till: [2021/12/02]

PPE to be used against category III risks





is in conformity with the provisions of Regulation (EU) 2016/425 and with the standards EN ISO 374-1:2016 Type B, EN ISO 374-5:2016, EN 420:2003 + A1:2009 and is identical to the PPE which is subject to the EU-Type examination (Module B, Annex V of the Regulation), under certificate number 032/2018/0719, issued by the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

and is subject to the procedure set out in Annex VII (Module C2) of the Regulation under the supervision of the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

Guido Van Duren Director - Regulatory

Director - Regulatory affairs

Ansell

Place: Brussels Date: 2018/04/19

The Manufacturer
ANSELL HEALTHCARE EUROPE N.V.
RIVERSIDE BUSINESS PARK, BLOCK J
BOULEVARD INTERNATIONAL 55
B-1070 BRUSSELS
BELGIUM

declares under his sole responsibility, that the PPE described hereafter:

### Neotouch 25-101

Products manufactured till: [2018/04/20]

PPE to be used against category III risks





is in conformity with the provisions of Regulation (EU) 2016/425 and with the standards, EN 420:2003 + A1:2009, EN 374:2003 and is identical to the PPE which is subject to the EC Type examination; under certificate number 03204121 issued by the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

and is subject to the procedure set out in Annex VII (Module C2) of the Regulation under the supervision of the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 70 B-9052 ZWIJNAARDE BELGIUM

Guido Van Duren

Director - Regulatory affairs

Ansell

Place: Brussels Date: 2004/06/01