

## Nitrile exam gloves with excellent dexterity along with enhanced comfort, fit and feel

- **Elevated comfort:** These thin, black nitrile medical exam gloves offer enhanced dexterity and tactility in a thin, yet strong design
- **Black nitrile:** Distinctive black color provides contrast while also hiding the appearance of oils, bodily fluids and other stains
- **ERGOFORM™ Technology:** ERGOFORM™ Ergonomic Design Technology reduces muscle strain during repetitive tasks, improving worker comfort and productivity
- **Optimized grip:** Textured on the fingertips for secure handling of instruments and materials



### Industries

- Food Service
- Cosmetics
- Tattoo
- Automotive Aftermarket
- Chemical
- Emergency Medical Services
- Healthcare
- Law Enforcement
- Life Sciences
- Dental
- Machinery and Equipment
- Metal fabrication

### Recommended For

- General dentistry
- Standard patient examinations
- Emergency services
- Catering/Meal preparation
- General Purpose Auto Aftermarket
- Equipment repair and maintenance
- Laboratory analysis
- Cosmetic Procedures
- Tattoo

# TECHNICAL DATA SHEET

Product Information	
External Glove Surface	Textured Fingers
Audit Standards	21 CFR 820, ISO 13485, ISO 9001
Product Certification	Medical Device Regulation (EU) 2017/745 Class I, Personal Protective Equipment Regulation (EU) 2016/425 Category III risks, EU Food Contact Approved, EC 2023/2006, Anvisa 81496190021 / CA 48.379, INVIMA 2015DM-0013417, CE 0493, EN 455 1-4, EN ISO 21420:2020, EN ISO 374-1:2016 Type B KPT, EN ISO 374-5:2016 Virus, EN 16523-1, ISO 18889:2019, ISO 9001, ISO 13485, UKCA, US QSR/FDA 510(k) Medical Examination Grade, FDA21 CFR 177-2600-US Food Contact Approved, ASTM D6319, EN 421:2010
Packaging Overview	Case=1000 gloves/10 dispensers; Dispenser=100 gloves 2XL: Case=900 gloves/10 dispensers; Dispenser=90 gloves
Storage Instructions	Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition.
Country Of Origin	China
Product Segmentation	High Touch
Antistatic	No
Silicone Free	No

## PHYSICAL PROPERTIES

	Typical Values		Testing Method
Length (mm/inches)	245 / 9.6		ASTM D6319, EN 455-2
Freedom from Holes (Inspection level I)	1.5 AQL		ASTM D5151, EN 455-1
Palm Thickness (mm/mils)	0.07 / 2.8		ASTM D6319, EN 455-2
Finger Thickness (mm/mils)	0.11 / 4.3		ASTM D6319, EN 455-2
	<b>BEFORE AGING</b>	<b>AFTER AGING</b>	
Ultimate Tensile Strength (MPa)	≥ 18	≥ 16	ASTM D412 & D573
Elongation at Break (%)	≥ 500	≥ 400	ASTM D412 & D573
Force at break (N)	≥ 6	≥ 6	EN 455

## ORDERING INFORMATION

Size	XS (5.5 - 6)	S (6.5 - 7)	M (7.5 - 8)	L (8.5 - 9)	XL (9.5 - 10)	XXL (10.5 - 11)
Product Code	93732060	93732070	93732080	93732090	93732100	93732110

## Performance Standards and Regulatory Compliance



## For additional information visit us at [www.ansell.com](http://www.ansell.com), or call us at

**Europe, Middle East & Africa Region**  
Ansell Healthcare Europe NV  
T: +32 (0) 2 528 74 00  
F: +32 (0) 2 528 74 01

**Asia Pacific Region**  
MY +603 8310 6688  
apac.medical@ansell.com  
CN: +86 21 38275000  
infochina@ansell.com  
JP: +813 5549 8151  
info.medical.jp@ansell.com

**North America Region**  
Ansell Healthcare Products LLC  
T: +1 800 800 0444  
F: +1 800 800 0445

**Latin America & Caribbean Region**  
BR +55 (11) 3356-3100  
latam.medical@ansell.com  
MX: +52 442 296 2050  
latam.medical@ansell.com  
CO: +57 1 288 3247  
latam.medical@ansell.com

**Australia**  
Ansell Limited  
T: +61 1800 337 041  
F: +61 1800 803 578

**Russia**  
Ansell PYC  
Ten. +7 495 258 13 16

## Technology



Ansell, ® and ™ are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending; [www.ansell.com/patentmarking](http://www.ansell.com/patentmarking) © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

